

I. Air velocity**I. 1) Calibrated Instrument**

Manufacturer: Swema
 Type: 3000MD
 Device Designation: Multifunction instrument
 Probe Designation: Hot wire probe and integrated temperature sensor
 Device Serial Number: 695579
 Probe Serial Number: 389839
 Device ID: 116205
 Probe ID: 116205

I. 2) Calibrating Conditions

Operator: Bjarke Kjær Olesen
 Date: 16-10-2023
 Ambient Temperature: 23,0 °C
 Relative humidity: 37 %HR
 Atmospheric Pressure: 1010 hPa
 Calibrating Principles: The point of calibration are realized with means of calibration according to : bench velocity WT180-500, measuring from 0,3 to 28 m/s controlled with flow transmitter CTV310, and two pressure transmitters CP301, CP303. All devices are held against the references instrument SWEMA 3000md sn: 676029 with probe SWA-31 sn: 386429 and pitot tube sn: 12972, which is traceable to national standard by SP certificates.

I. 3) Measurement results

Vr (m/s)	Vi (m/s)	Vi-Vr (m/s)	Uncertainty (m/s)
0,3	0,26	-0,040	0,04
0,6	0,59	-0,010	0,04
1	0,99	-0,010	0,05
2	1,96	-0,040	0,07
3	2,94	-0,060	0,08
4	4,0	0,000	0,10
6	6,0	0,000	0,11
8	7,9	-0,100	0,13
10	10,0	0,000	0,14
15	15,3	0,300	0,17

Vr: value displayed by our reference instrument, Vi: valeur displayed by customer's instrument.

Comments: The temperature checking give: Vr 25,0°C ; Vi 25,0°C

the measurement uncertainty stated is a combination of laboratory and shot term contributions from calibration item. The uncertainty is given with a coverage factor of 2 corresponding to a coverage probability of approx. 95%.
 the values stated are valid for the calibration item only and are mean values of 3 successive readings. The measurement uncertainty is calculated in accordance with EA-4/02.